



Illumina, Inc
25861 Industrial Boulevard
Hayward, CA 94545-2991
Tel: 510.670.9300
Fax: 510.670.9302
www.illumina.com

Certificate of Conformance

Manufacturer:	Illumina, Inc.
Name:	MiSeqDx Sequencing System
Model/Part Number:	MiSeqDx Instrument / 20014054
Serial Number:	C70289
Date of Manufacture:	2020-01-28
RTA Version:	1.18.54.3
MOS Version:	2.2.4.3
LRM Version:	1.4.5.0

This document certifies that each instrument conforms to Illumina specifications and is developed and validated following a certified Quality System, which conforms to ISO 13485 requirements. The integrity of the Illumina Quality System is routinely audited and is certified by the British Standards Institute (BSI, Inc.), a quality management system registrar, as meeting the requirements of ISO 13485.

Each instrument goes through a sequencing run in the Final Instrument Test with defined acceptance criteria. In accordance with the Device History Record, each instrument must pass the Final Instrument Test and cosmetic and labeling verifications, which are performed by qualified and trained manufacturing personnel. The Device History Record is then approved by independent, qualified Quality Assurance personnel.

Verified by:

Angela Murray
Angela Murray, Quality Ops Supervisor

Date:

10-FEB-2020